

HSCC(I)Ltd.

Amendment -XV

dated 31.10.2017

Tender Enquiry No. : HSCC/KCGMC/Medical Equipment/2015-06/14 dated: 30/06/2017

Bid sale, submission and opening date for items 2 & 5 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	10.11.17, 3.00 P.M.
ii.	Closing date & time for receipt of tender	10.11.17, 3.30 P.M.
iii.	Time and date of Opening of Tenders	10.11.17, 4.00 P.M.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

Following amendments may be considered in Laproscope.

1. Full HD Digital Camera
 - A. The System should have digital and integrated / C - mount optical zoom to enhance quality of image and cross speciality standardization of camera system, regardless of telescope used.
 - B. Camera head weight should be less than 500 grams.

Rest all specifications remain the same.

2. Laproscope / telescope length for
 - 10 mm may be amended to 30 cm or more.
 - 5mm may be amended to 27 cm or more.Rest all specifications remain the same.

3. Grasping forceps may be amended as
 - Should be dismantable into two / three parts
 - Length between 33 - 36 cms.

s/d

CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula