

AMENDMENT No – IV Dated 05/03/2016

Sub.: Procurement of Medical Equipment for Kalpana Chawala Government Medical College, Karnal.

Ref: Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015-05 dt. 18.01.2016

Bidder is requested to note the following Amendments:

1. DEPARTMENT OF OPHTHALMOLOGY

Sl. No.	Name of the Article	EMD (Rs)	Revised EMD (Rs)
2	Operation Theatre Table Hydraulic for Ophthalmology	1,40,000/-	10,000/-

2. DEPARTMENT OF ENT SCHEDULE 1: - ENT Treatment Unit

S. No.	Specification	Amended Specification
•	Endoscopy center	Should be integrated
•	Patient examination chair	
5	Foot pedal and hand switch	Foot Pedal or Hand Switch
•	Diagnostic Microscope	
3	Magnification changer with 20-30cm focal length	Magnification changer with 15 cm - 30 cm focal length
5	3 step magnification changer	Stands Delete
	Certification	Should be USFDA or European CE approved

SCHEDULE 2: - Direct Laryngoscopy set

S. No.	Specification	Amended Specification
1	Anterior Commissure Laryngoscope : length 17 cm medium, universal size, triangular spatula shaped, lateral outer channels for fibre optic light carrier and suction, with FO light carrier for laryngoscope	Anterior Commissure Laryngoscope : length 17/18 cm medium, universal size, triangular spatula shaped, lateral outer channels for fibre optic light carrier and suction, with FO light carrier for laryngoscope
3	Negus Laryngoscope : operating laryngoscope for children, 15 cm with FO light carrier for distal illumination and suction tube to remove vapour	Operating laryngoscope for children, 15-18.2 cm with FO light carrier for distal illumination and suction tube to remove vapour
6	Laryngeal Suctions 2mm, 4mm, 6mm, 23 cm	Laryngeal Suctions 2mm, 3mm, 23 cm
	Certification	Should be USFDA or European CE approved

SCHEDULE 3: - Microlaryngoscopy set

S. No.	Specification	Amended Specification
•	Kleinsassers Laryngoscope Adult, 18cm large or medium,	Kleinsassers Laryngoscope Adult, 18/18.2 cm large or medium,
•	Suspension system	Laryngoscope holder and chest support
	Certification	Should be USFDA or European CE approved

SCHEDULE 4: - Mastoidectomy Set

S. No	Specification	Amended Specification		
•	Electric Drill (Motor Hand piece & burr)	Electric Drill (Motor Hand piece & burr) <ul style="list-style-type: none"> • Speed of drill system should be at least 50,000 rpm or above • Double pedal footswitch to control speed • hand pieces: straight (1) & angled (1) 		
	2mm	2mm	Cutting Diamond (10 each)	Diamond (5 each)
	3mm	3mm	Cutting Diamond (10 each)	Diamond (5 each)
	4 mm	4 mm	Cutting Diamond (10 each)	Diamond (5 each)
	5 mm	5 mm	Cutting Diamond (10 each)	Diamond (5 each)
•	Mastoid Seeker	Mastoid seeker/ Gross ear hook spoon		
•	Aditus seeker	Aditus seeker/ gross ear hook curette		
	Certification	Should be USFDA or European CE approved		

SCHEDULE 5: - Direct laryngoscopy set

S. No.	Specification	Amended Specification
2	Negus	Deleted
3	Fibre optic lighting system with light source	Fibre optic lighting system with light source of 300w xenon light source

SCHEDULE 7: - Brain stem Evoked Response Audiometer with ASSR

S. No.	Specification	Amended Specification
1	Upgraded with OAE, ASSR & VNG, NCT	Upgradable with OAE, ASSR & VNG, NCT

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will be published in the news papers in this regard.

CGM, HSCC India Limited
For and on behalf of DGMER, Panchkula