

## Amendment IV

HSCC/SJH/MED.Equt./2016

Dated: 27.06.2016

**Ref: Tender No. HSCC/SJH/Med. Eqpt./2015/18 dt. 18.05.2016**

**Subject: Procurement of Medical Equipment for New Emergency Block& Super-Specialty Block at Safderjung Hospital, New Delhi**

Due to inadequate response of bid submission date for the item No. 2, 3, 8, 10, 11,14,16,17, is extended to 08.07.2016

**Technical Amendment: -**

**Item No: 8 Laparoscope Sets**

**Existing as:**

LED Light Source 300 watts or more instead of 170 watts.

**Amended as:**

LED light Source 170 watts or more with LED lamp life of not less than 25,000 hrs. Or Xenon light source 300 watt with at least 5 no. of spare Xenon lamp 300 watt.

All other tender terms and conditions remain unchanged

Amendment to be issued will be uploaded on website [www. Tenderwizard.com/HSCC](http://www.Tenderwizard.com/HSCC) & [www.hsscLtd.com](http://www.hsscLtd.com)

**Medical Superintendent**  
**Safderjung**  
**Hospital & VMHC, New Delhi.**