## HSCC (India) Ltd

## <u>AMENDMENT –X</u>

Dated: - 29.09.2016

**Subject:** Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/ Medical Equipment / 2015-06/04 Dated 25.06.2016

Bid sale, submission and opening date for item no. 3, 4 has been extended as per details given in Table -1:

Table -1

SI. No.	Description	Revised Schedule
i.	Sale date of the tender	06.10.2016, 2.00 P.M.
ii.	Closing date & time for receipt of tender	06.10.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	06.10.2016, 3.00 P.M.

## Bidders are requested to note the following technical Amendment: Item No. 4- C-arm image intensifier

S. No.	Specification in the bid	Amended as
	5.)X- ray generator	5.)X- ray generator
	High frequency 2.5 KW or even better X-ray generator with high capacity rotating anode X-ray tube of dual foci.	High frequency 3 KW or even better X-ray generator with high capacity rotating/ stationary anode X-ray tube of dual foci.
	The complete functional system must be quoted with DUAL CHANNEL Laser LIGHT SOURCE ON, X-RAY TUBE UNIT for making a cross to reduce the X-ray dose, Built in DOSE AREA PRODUCT meter for disply of X-ray dose, light weight lead aprons (6) required CVT and thermal imaging film printer with 5 film rolls WITH DRY CHEMISTRY CAMERA & 500 DPI FOR HARD COPY DOCUMENTATION	The complete functional system must be quoted with DUAL CHANNEL Laser LIGHT SOURCE ON, X-RAY TUBE UNIT for making a cross to reduce the X-ray dose, Built in DOSE AREA PRODUCT meter for display of X-ray dose, light weight lead aprons (6) required CVT

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

CGM, HSCC India Limited For and on behalf of DGMER, Panchkula