

Amendment -V dated 07.09.2016

IFB: HSCC/KCGMC/Medical Equipment /2015-03/C dated 27.06.2016

All bidders are requested to note the following:

Item no. & Name	Tender Technical Specification	May please read as
Item no. 03, Upper GI Endoscope	Main Lamp: Xenon Short-arc Lamp (300 Watts) with Switching Regulator Mechanism	100 watts or more
	Emergency Lamp: Halogen 12V 100 Watts / LED with apprx 500hrs or more life. (with Two spare bulbs).	Emergency lamps: Halogen/ LED/Xenon
	Output: RGB, Y/C, Composite (with Simultaneous Output Possibility)	Output: of DVI,RGB, Y/C, composite (In camera control unit)
Item no. 05 CO2 YAGLASER		Added Scanner hand piece
Item no. 06, CONTACT MICROHYSTROSCO PE	BETTOCCHI Operating Sheath, size 4.3mm, with channel for semirigid 5 Fr. operating instruments, with 1 stopcock and 1 LUER-Lock adaptor, for use with CF operating sheath – 1 No	BETTOCCHI deleted
	BETTOCCHI Continuous-Flow Operating Sheath, size 5.mm, with 1 stopcock and 1 LUER-Lock adaptor, for use with operating sheath – 1 No	BETTOCCHI deleted

Item no.	Last date & time on line sale of Tender enquiry document	Last closing date & time for receipt of e-tender	Last Time and date of opening of Techno – Commercial bids.	Earlier date of opening
1 to 8	16.09.2016, upto 12.30 hrs IST	16.09.2016, 13.00 hrs IST	16.09.2016, 13.30 hrs IST	09.09.2016

As per SCC clause no. 9, the validity of bid security (EMD) shall be considered from the original date of submission of bid/ tender in case tender opening date extended.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d

CGM, HSCC (I) Ltd.
For & on behalf of DGMER, Panchkula