

HSCC (India) Ltd
PROCUREMENT
Amendment – IX

Dated: 05.05.2017

Subject: Amendment of e- Tender for Kalpana Chawla Govt. Medical College, Karnal..
Ref.IFB: HSCC/KCGMC/Medical Equipment /2016-17/MC-02 dt 20.02.2017

1. The Schedule of Tender remains as under:

Sl. No.	Description	Schedule
1	Closing date & time for receipt of Tender	18.05.2017, 14:30 Hrs
2	Time and date of opening of tenders	18.05.2017, 15:00Hrs

2. Specification and EMD Amendment:

A) Following should be read along with the specifications for Glassware already provided for Pathology, Physiology, Biochemistry & Microbiology Departments ONLY:

1. GLASSWARE DETAILS:

Glassware Quality:

All glassware items are made from borosilicate 3.3 low expansion glass tube.

Salient Features of Glassware Items :

- Uniform wall thickness
- Accuracy in calibration, as per Certificates – A or B types as required
- Thermal shock resistant
- Enhanced Mechanical Strength
- Chemically resistant to attack from acids organic solvent.

Certifications :

Principal Manufacturer should have valid NABL, CE, GMP, ISO 9001 and ISO 13000 quality Certificates.

The Glassware should of any of the following makes only:

- | | | |
|-------------|-----------|------------|
| 1. Borosil | 3. Pyrex | 5. Corning |
| 2. Supertek | 4. Fisher | |

B) The EMD of the Glassware for the following Department stands AMENDED as follows:

I. Department of Pathology, KCGMC, Karnal

	NAME OF EQUIPMENT	QTY (Nos)	EMD INR (Rs.)
11	GLASSWARE	Various Qty	2,000.00

All other terms and conditions of the IFB document remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP website for corrigendum/amendments etc. if any, as no separate Advertisement / correspondences shall be made.

CHIEF GENERAL MANAGER, HSCC (I) Ltd.
For & on behalf of DGMER, Panchkula.