

AMENDMENT - V**Dated: 20.05.2014****INVITATION FOR BIDS FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF MEDICAL EQUIPMENTS FOR 200 BEDDED NEPAL- BHARAT MAITRI EMERGENCY AND TRAUMA CENTRE (NBMETC), KATHMANDU, NEPAL****Bid No.: BID REFERENCE: IFB No. HSCC/PUR/NEPAL/2013-14/2 Dated 04.03.2014**

Bidder is requested to note the following amendments:

Amendment – Technical**Item no: 8 for Ultrasound Machine**

Sr. No.	Existing Specification	Amended Specification
Point no: 10 & 11	10 It should be possible to upgradeable the system to 4D, later if required. Please confirm. 11 System should be able to do volume acquisition of 33 volumes/sec or more.	System should be able to do volume acquisition of 30 volumes/sec or more.
Point no: 14	The system should have imaging depth of 25 cm or more.	The system should have imaging depth of 30 cm or more
Point no: 18	The system should have an easy to use control panel and also be rotated sidwise more than 180 degrees. Should have an alphanumeric keyboard with illuminated keys and status display.	The system should have an easy to use control panel and also be rotated sidwise more than 100 degrees. Should have an alphanumeric keyboard with illuminated keys and status display and should have height adjustable control panel.
<u>Transducers:</u>	4. Small foot print, phased array transducer for Transfontanelle USG of neonatal head.	Deleted
Additional Requirement : Certification		US FDA approval/ Certification required

Item no.: 10 High Speed Refrigerated Centrifuge

Sr. No.	Existing Specification	Amended Specification
Point no: 1	Floor model cold centrifuge with temperature setting in a range of -4°C to +40°C with CPC free condenser.	Floor model cold centrifuge with temperature setting in a range of -4°C to +40°C with CFC free condenser.
Point no: 4	Access to various angles rotors with variable capacity to adopt 250ml x 4, 120ml x 4, 50ml x 10, 15ml x10-12, 1.5mlx 18-24 and 96 well microtitre plates.	Access to various Fixed angles rotors with variable capacity to adopt 250ml x 4, 100-120ml x 4, 50ml x 10, 15ml x10-12, 1.5mlx 18-24 and 96 well microtitre plates.

Other technical specifications remain unchanged

Amendment – Commercial

Item no.1,2,3,4,5,6,7 is deleted from the requirement.

Bidder is requested to note the following amendment in the bid sale, submission and opening date of the item no. 8-12 of the aforesaid bid:

S. No.	Details	Revised schedule
1.	Last date and time for the sale of bid	29.05.2014
2.	Last date & time for submission of bid	30.05.2014 by 14.30 hrs.
3.	Techno-commercial opening of bids	30.05.2014 by 15.00 hrs.
4.	Venue of bid opening	HSCC (India) Ltd., E--6A, Block-E, Sector-1, NODA (U.P.)

All other terms and conditions of the bid document shall remain unchanged.

S/d
Joint Secretary –DPA (III)
Ministry of External Affairs,
New Delhi