

**ALL BIDDERS****Amendment-1**

**Subject: Supply, installation, testing & commissioning of Modular Type Office & Laboratory furniture & associated works etc for Regional Medical Research Centre at Dibrugarh, Assam**

Dear Sir,

This has reference to the pre bid meeting held on 19.01.2012 at 15:00 hrs for the subject work at HSCC corporate office, Noida,

The following AMMENDMENT-1/CLARIFICATION-1 may be noted, which shall be treated as a part of the contract to be submitted duly signed & stamp along with tender/contract **to be submitted duly signed & stamp along with tender envelope no. - 2:**

Sl. no.	Queries raised by bidders	Clarifications
1.	The floor drawings provided does not have any measurement.	As per tender conditions, site may please be visited.
2.	There was no furniture layout provided in any of the annexures.	Tender drawings indicating tentative layout of furnitures is attached.
3	In the absence of furniture layout, it is very difficult for the bidders to understand.....	As per tender drawings, layout of furnitures to be submitted by agency for approval before execution.
4	Fume hoods specified in the tender lack certification compliance to ASHRAE-110 and EN -14175. Fume hoods not confirming to these performance standards might pose serious health hazards to the end users later	As per tender conditions only
5	The tender specifies the fume hood to have as exhaust as optional. However please note that without ducting/blower, the fume hood would be non-functional. Moreover, ducting material specified in FPT whereas recommended material is PP-FRP (3+3 mm thick) which is rigid, temperature resistant and of superior quality. The tender also lack the	In item no 43 of BOQ ( Vol-IV) FPT may be read as FRP. Ducting route shall be finalized before execution however site may please be visited to get the tentative route of ducting.

	appropriate drawings for designing optimum ducting route and the approved blowers.	
6	The laboratory furniture specified in the tender lack WELS certification. The specification enables the building/facility to be labeled under Green....	As per tender conditions only
7	The Laboratory furniture technical specifications suggest a design which is out dated and no longer.....	As per tender conditions only
8	There is, as far as our knowledge goes, no mention about the road permits and tax forms in the tender conditions,. Will the road permits and the form C will be provided for the ....	As per tender conditions only
9.	Specification of Modular furniture is not well laid..... .	As per tender conditions only
10	All the workstations asked for, in the tender is fitted with Drawers, KBPT & CPU trolley. Over and above these there are 100 nos. each of drawers, KBPT & CPU trolley is asked. Where these items will be installed.	The revised quantity of following BOQ items may be read as follows: <ul style="list-style-type: none"> <li>- BOQ item no. 11 : <b>Pedestal Drawer Unit (2+1)</b> : quantity may be read as <b>09</b> in place of <b>100</b></li> <li>- BOQ item no. 12 : <b>Key Board Tray</b> : quantity may be read as <b>09</b> in place of <b>100</b></li> <li>- BOQ item no. 13 : <b>CPU Trolley</b> : quantity may be read as <b>09</b> in place of <b>100</b></li> </ul>
11	Time limit as mentioned in the tender document is 02 weeks, which is unrealistic.....	As per tender conditions only

Other conditions of tender will remain unchanged.

Thanking you,

Yours faithfully,

(S. Mukhopadhyay)  
DGM (Civil)