

All Bidders**Amendment -I**

Subject: Supply, Installation, Testing & Commissioning of Medical Gas Manifold System along with buy back of existing manifold gas pipeline as on where is basis at CT-2 CCU, AIIMS, New Delhi.

IFB No. : HSCC/PG-II/AIIMS-CCU/MGMS/2013

Dear Sir,

This has reference to above IFB No. for the Subject works.

The following Amendment may be noted which shall be treated as part of the contract to be submitted duly signed & stamp along with tender.

	Tender Conditions/Bidder's Query	Amendment/Clarification
1.	Flow meter with Humidifier bottle at Sr. No. 7 (g). As per tender requirement 'The Humidifier bottle is made of unbreakable & reusable of polycarbonate materials autoclavable 121 deg C'. It is suggested that as per American Standard the polypropylene should be added to polycarbonate.	May be read as The Humidifier bottle is made of unbreakable & reusable of polycarbonate materials/ polypropylene autoclavable 121 deg C.
2.	HP & LP Tubing HP & LP Tubing Flexible having Antistatic core a per ISO with proper colour coded.	Colour code is on HP tubing only.
3	In item ward vaccum units (Imported) Collection bottle size – 2000 ml	Amended as Collection bottle size of 1700ml -2000 ml.

Bidder should follow the tender terms & condition for the unanswered queries.

All other terms & conditions remain unchanged.

CGM (D&E & Proj.)