## Amendment – II

## Bid No.: IFB Reference: HSCC/PUR/MEA - Myanmar/Equipment/2013 -14 dated 6<sup>th</sup> May 2013

Bidders are requested to note the following amendment:

## Package 4 – BLS Ambulance

SI No.	Existing Tender Specification	Amended by MEA/HSCC
1	Annexure - I	No Change
	The Ambulance shall be left hand drive.	
2	Annexure – V (D) STRETCHER	
	<ol> <li>One scoop stretcher, aluminum makes which should include 3-4 restraints straps.</li> <li>Combi Cot Stretcher</li> </ol>	<ol> <li>One scoop stretcher, aluminum makes which should include 3-4 restraints straps CE/ FDA compliant.</li> <li>Combi Cot Stretcher CE/ FDA compliant.</li> </ol>
3	(E) AIR CONDITIONER AND HEATING	
	HEATING	
	1. Separate 220V heating blower should be provided for patient cabin temperature.	No Change
4	Annexure - VII	No Change
	A Suction Devices battery operated	
	<ol> <li>An engine vacuum operated or electrically powered and battery backup (in built}, complete suction aspiration system. Shall be installed permanently on board to provide for the primary patient. It shall have wide bore tubing.</li> </ol>	
	2. A manual suction device, age and weight appropriate, with wide bore tubing and at least a six ounce reservoir.	
	There must be an assortment of suction (minimum of 2 each) on board. Sizes 6 fr, 8fr, 10fr, 16fr, 18fr. A rigid suction catheter (e.g. Yankaur) will also be carried. Minimum 2 each.	
	Annexure – VII (P -1) P. Spinal immobilization devices: Cervical collars to accommodate the infant.	

	Child. Medium adult and large adult	Deleted
	sizes. Collars must be manufactured of semi-	
	rigid or rigid material.	
5	Electrical system with power backup	No Change
6	Annexure - VIII Additional Equipment	No Change
7	Warranty, EMD, Performance Security	As per the bid

## Package -5 Operation Theatre Light - OT sky lamp (Twin)

SI No.	Existing Tender Specification	Amended by MEA/HSCC
2	Point 3.1 (9) : Illuminated field diameter range adjustable approx 10-15 Cm.	Illuminated field diameter range adjustable by approx 10-15 Cm i.e if at a particular distance the min illuminated field diameter is X, it should adjustable to X+10-15 Cm

All other terms and conditions of the bid document shall remain unchanged.

Joint Secretary –DPA (III)

Ministry of External Affairs,

New Delhi