

October 25, 2012

“REPLY TO BIDDERS QUERIES AND AMENDMENT TO TENDER”:-

**Sub. Supply Installation Testing & Commissioning of Oxygen Concentrator Module & Medical Gas Manifold System at Yangon Children Hospital, Myanmar**

**&**

**Oxygen Concentrator Module with Pipeline system for 4 OTs at Sittwe General Hospital, Myanmar**

. Tender No : HSCC/PUR/MEA-Myanmar/OCM/2012

Date : 01.10.2012

**CLARIFICATION / AMENDMENT No. – (III)**

<b>Previous amendment as per Amendment-II</b>	<b>Amended as</b>
Compressed Air System consisting of Compressor of non-lubricated type to meet the peak atmospheric air and pressure requirement Sr. 1.2 of Yangon Children hospital and Sittwe General Hospital	Compressed Air System consisting of Compressor (Non-lubricated type) and Refrigerated Air Drier of Sr.No.1.2 for Yangon Children hospital and Sittwe General Hospital.
Compressed Air System consisting of Compressor of non-lubricated type to meet the peak atmospheric air and pressure requirement Sr. 1.2 of Yangon Children hospital and Sittwe General Hospital	The Oxygen Concentrator should be supplied with Air Compressor (Non lubricated type) System to meet the peak atmospheric air and pressure requirement.

All other terms and conditions of the tender documents shall remain same.

**Jt. Secretary (DPA III)**

**MEA, New Delhi**