

AMENDMENT-I

Ref: Tender Enquiry No.: HSCC/SJH/Med. Eqpt. /36 dated 5.11.2017

Sub.: Procurement of Medical Equipment for New Emergency Block & Super- Specialty Block at Safderjung Hospital, New Delhi.

Amendments have been received Safdarjung Hospital for Item No. 1 Gamma Knife (Proprietary), Item No. 2 Mobile CT Scan & Item No. 3 Hematology Analyzer It is therefore extended the bid submission date from 28.11.2017 to 05.12.2017.

Amendments are as follows:-

Item No. 1 GAMMA KNIFE SYSTEM WITH ADVANCE IMAGE GUIDANCE AND ADAPTIVE DOSE CONTROL FOR INTRA-CRANIAL STEREOTACTIC (Propriety item)

| Sr. No. | Bidder Queries | Tender Specification | Amendment |
|--------------|--|---|---|
| Point No.22 | Kindly Amend this to CBCT Imaging Accuracy:≤0.5mm | CBCT Imaging Accuracy:≤0.05mm | CBCT Imaging Accuracy:≤0.5mm |
| Point No. 46 | Kindly Amend this to Should support at least 50 isocenters per plan. | Should support at least 500 isocenters per plan | Should support at least 50 isocenters per plan. |
| Point No. 59 | Kindly Delete this specification as in latest model of Gamma knife this feature is not required due to new added feature of mask based multi fractionation which fulfills same purpose with more ease and comfort to patient to treat large tumor volumes. | It should have a non-invasive repeat fixation device based on vacuum assisted mouthpiece and should be fully integrated with the stereotactic robotic system and should enable fractionation. | Delete this specification as in latest model of Gamma knife this feature is not required due to new added feature of mask based multi fractionation which fulfills same purpose with more ease and comfort to patient to treat large tumor volumes. |

Note: Layout Plant will be uploaded shortly.

Item No. 2

MOBILE CT SCANNER FOR NEUROSURGERY / TRAUMA ESSENTIAL ACCESSORIES FOR USE IN OPERATION THEATER

| Sr. no | Bidder Queries | Tender Specification | Amendment |
|--------|--|------------------------|--|
| 10 | Not Required as system is hooked to PACS for data storage and further processing. (image from work Station can be pushed directly to PACS and further processed through PACS to achieves portability of CT | X-RAY FIELD DIGITIZER: | Point Deleted as system is hooked to PACS for data storage and further processing. |

| | | | |
|----|--|--|----------------------|
| 11 | Option has been removed from care tomodoro headrest system. Instead OT table adaptor is provided | HORSE SHOE HEAD REST (ADULT & PAEDIATRIC - 1 EACH) | Point stands Deleted |
|----|--|--|----------------------|

Item No. 3

Hematology Analyzer

| <u>Sr. No.</u> | <u>Bidders Queries</u> | <u>Tender Specification</u> | <u>Amendments</u> |
|-----------------------|--|---|--|
| 4 | The instrument should have throughput of at least 60 samples / hr. in CBC+Diff mode. | The instrument should have throughput of at least more than 75 samples per hour in CBC+DIFF MODE. | The instrument should have throughput of at least more than 60 samples per hour in CBC+DIFF MODE. |
| 16 | Equipment, reagents and reportable parameters should be US-FDA plus CE approved. | Equipment, reagents and reportable parameters should be US-FDA/EU-CE approved. | Equipment, reagents and reportable parameters should be US-FDA & CE approved. |
| 26 | | Start up kit for 5000 tests and 02 packs of control (Higher, Low, Normal) along with equipment at the time of installation. | As advised by Client: Start up kit for 5000 tests of CBC+DIFF MODE and 02 packs of control (Higher, Low, Normal) along with equipment at the time of installation. |

Amendment to be issued will be uploaded on websites www.tenderwizard.com/HSCC & www.hsccltd.com.

All other tender terms and conditions remain unchanged.

Medical Superintendent
Safdarjung Hospital VMMC,
New Delhi

