

Amendment no 01 Dated 31.05.2016
IFB no. HSCC/KCGMC/Medical Equipment/2015-03/B dated 05.05.2016

Item no. & name of items	Tender Specification	Tender Specification may please read as
Item no. ECG Machine 01 Channels	<p>Point No. 4:- The unit should be lightweight and portable, capable of storing at least 100 or 200 electrocardiograms and allow for rapid transfer of this data to a computer through wireless or USB ports. DICOM/HL7 ready for data transfer to a HIS/PACS. Should be possible to collate ECGs based on data like ECGs with QRSd>120ms or QT > 450ms</p> <p>Point No. 5:- Machine should run both on AC current & built in rechargeable batteries and should be provided with a paper bat to carry the equipment and its ancillaries. A moveable original CART to which the machine can be fixed has to be supplied in addition.</p> <p>Point No. 7:- The unit should have the capability for proper self-check & calibration.</p>	<p>Point No. 4:- The unit should be lightweight and portable, capable of storing at least 100 or 200 electrocardiograms and allow for rapid transfer of this data to a computer through LAN/USB ports. DICOM/HL7 ready for data transfer to a HIS/PACS. Should be possible to collate ECGs based on data like ECGs with QRSd>120ms or QT > 450ms</p> <p>Point No. 5:- Machine should run both on AC current & built in rechargeable batteries and should be provided with a paper bat to carry the equipment and its ancillaries. A moveable CART powder quoted with basket to which the machine can be fixed has to be supplied in addition.</p> <p>Point No. 7:- The unit should have the capability for proper self-check.</p> <p>The product should be USFDA / European CE approved.</p>

Item no.	Last date & time on line sale of Tender enquiry document	Last closing date & time for receipt of e-tender	Last Time and date of opening of Techno – Commercial bids.	Earlier date on opening
1, 2 & 3	09.06.2016, upto 12.30 hrs IST	09.06.2016, 13.00 hrs IST	09.06.2016, 13.30 hrs IST	02.06.2016

As per SCC clause no. 9, the validity of bid security (EMD) shall be considered from the original date of submission of bid/ tender in case tender opening date extended.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will published in the news papers.

s/d

CGM, HSCC (I) Ltd.
For & on behalf of DGMR, Panchkula