

# HSCC (India) Ltd

## AMENDMENT – I

Dated:- 12.03.2016

**Subject:** Amendment to the tender Enquiry Document.

**Ref: Tender Enquiry No.:** HSCC/KCGMC/ Medical Equipment / 2015-06/01 Dated 22.02.2016

Bidders are requested to note the following amendments:

### Technical Amendments

#### Item No. 1. Flexible Endoscope, Upper GI & Colonoscope

S.No.	Description	Amended Specification
<b>A</b>	<b>Video Gastro-scope – Should have the following:</b>	
	Capable of producing Endoscopy images like NBI/SPIES/I-SCAN, with following features:	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc.
<b>B</b>	<b>Video Colonoscope – Should have the following:</b>	
	Capable of producing Endoscopy images like NBI/SPIES/I-SCAN, with following features:	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc.
	Equipment cart rides on 4 antistatic wheels, equipped with atleast 02 locking brakes, 3 fixed shelves, 1 with handles and lockable drawer, scope hanger to mount flexible scope.	Equipment cart rides on 4 antistatic wheels, equipped with atleast 02 locking brakes, 3 fixed shelves, 1 with handles and drawer, scope hanger to mount flexible scope.

## HSCC (India) Ltd

### Item No. 3 Diagnostic Laparoscope (2 Nos.) & Operative Laparoscope (2 Nos)

S.No.	Description	Amended Specification
•	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better. Camera & Coupler should be one piece (Integrated)	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better.
•	LED light source 220volts, 225-300watts OR Xenon light source 300 watts or more	Xenon Cold light fountain with 300 watts or more
•	Should have more than 5000 to 25000 hrs bulb life (with one or spare bulbs)	Xenon Lamp life 500 Hours or more (With 2 xenon spare bulbs)

### Item No. 4. Bronchoscope Video Flexible

S.No.	Description	Amended Specification
5	Outer diameter should be 5.8 – 6.3 mm.	Outer diameter should be 6.0 mm or less

### Item No. 6. Paediatrics Video Flexible Sigmoidoscope

S.No.	Description	Amended Specification
3	It should have 150watts Halogen lamp	It should be amended 150 watts or more
5	Distal End Diameter should be minimum 13.4mm.	Distal End Diameter should be minimum 13mm / 13.4mm.
6	Insertion Tube Diameter should be 12.8mm	Insertion Tube Diameter should be 12.8mm / 13.2mm

## HSCC (India) Ltd

9	It should have a rotatable PVE Connector for easier maneuverability by 180 Degree	It should have a rotatable PVE Connector for easier maneuverability by 180 Degree or equivalent
---	---	---

### **Item No. 09. Ultrasound (Coloured) (Radiology - 1 No, Surgery - 2 Nos, Medicine - 1 No &Obs/Gyn - 5 Nos)**

S.No.	Description	Amended Specification
Reference	Tender Specs	
Point 3	The system should have fully digital technology with minimum 20000 digital processing channels per image frame for simultaneous formation, acquisition and display processing of multiple ultrasound beams and support dynamic focal length tuning.	The system should have fully digital technology with minimum 1,50,000 digital processing channels per image frame for simultaneous formation, acquisition & display processing of multiple ultrasound beams & support dynamic focal length tuning
Point 5	A display of 17" or more high resolution LED/LCD Flat Panel/High Resolution Screen with swivel and tilt facility. It should also have a touch panel on the console for ease of operation of minimum size 7"	A display of 19" or more high resolution LED/LCD Flat Panel/High Resolution Screen with swivel and tilt facility. It should also have a touch panel on the console for ease of operation of minimum size 7"
Point6	Inbuilt image storage facility with not less than 320 GB HDD and DVD Writer facility. The image management must enable to rework on the volume files that are stored in the HDD.	Inbuilt image storage facility with not less than 320 GB HDD or more and DVD Writer facility. The image management must enable to rework on the volume files that are stored in the HDD.
Point 10	Transmission focus must be freely selectable in 1 to 5 focal zones and adjustable in minimum 6 different positions.	Transmission focus must be freely selectable in 1 to 4 focal zones and adjustable in minimum 6 different positions.
Point22	"The system should have a Battery backup of at least 90 minutes for uninterrupted continuous scanning".	"The system should have a Battery backup or UPS of at least 90 minutes for uninterrupted continuous scanning".

## HSCC (India) Ltd

Point 26	Safety conformance: Should meet all the standard norms & it should be ISO, CE /US FDA approved.	Should be US FDA & European CE
Point 29	Manufacturer should be US FDA or European CE Certified/Approved	Should be US FDA & European CE
		Price of Additional probes may be quoted as optional for Paediatric &/ Cardiac use, Trans Oesophageal Probes

**Bid Security submitted in the form of Bank Guarantee should be in the name of HSCC I Ltd., E-6(A), Sector -1 Noida**

**Bid sale, submission and opening date for following items has been extended as per details given in Table -1:**

**Table -1**

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	21.03.2016, 2.00 P.M
ii.	Closing date & time for receipt of tender	21.03.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	21.03.2016, 3.00 P.M.

- **Addition of SCC Clause :1 in the bid document as per following:**

**Reimbursement of Custom Duty:** The custom duty amount as mentioned in the price schedule section –XI (B) (convert in INR at the rate of exchange mentioned in bill of entry) will compared with the actual total custom duty amount paid to custom department and the same will be reimbursed to the supplier as per following:

- If the custom duty amount as mentioned in the price schedule section –XI (B) is equal to the actual total custom duty amount levied by custom department, the actual total custom duty amount levied by custom department shall be prevailed and reimbursed to the supplier in INR accordingly.
- If the custom duty amount as mentioned in the price schedule section –XI (B) is more than actual total custom duty amount levied by custom department, the actual total custom duty amount levied by custom department shall prevail and reimbursed to the supplier in INR accordingly.

## HSCC (India) Ltd

---

- c. If the custom duty amount as mentioned in the price schedule section –XI (B) is less than the actual total custom duty amount levied by custom department, the custom duty amount as mentioned in the price schedule section –XI (B) shall prevail and reimbursed to the supplier at rate of exchange rate mentioned on the bill of entry in INR accordingly.
- **The Price schedule for the Goods to be imported from abroad Section XI – B has been modified and attached at Appendix –A with this Amendment.**

**Bid Document Fee:** Micro and small scale enterprises (MSEs) are exempted from Bid Document fee, The MSEs fulfilling the prescribed eligibility criteria and participating in the tender shall enclose with their tender a copy of their valid registration certificate with District Industries Centres or Khadi and Village Industries Commission or Khadi and Village Industries Board or Coir board or National Small Industries Corporation or any other body specified by Ministry of Micro and Small enterprises in support of their being on MSE, failing which their tender will be liable to be ignored.

**All other terms and conditions of the tender enquiry document shall remain unchanged.**

**Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.**

**CGM, HSCC India Limited  
For and on behalf of DGMR,  
Panchkula**

# HSCC (India) Ltd

Appendix - A

## SECTION – XI B) PRICE SCHEDULE

### PRICE SCHEDULE FOR GOODS TO BE IMPORTED FROM ABROAD

1	2	3	4	5					6
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Price per unit (Currency)					Total price on CIP Port of destination + Extended Insurance+ local transportation and storage at consignee site)  4X 5 (f)
				FOB/FCA price at port/ airport of Lading  (a)	Carriage & Insurance (port of loading to port of destination) and other Incidental costs  (b)	CIP Price (name place/port of destination in India  (c)= (a) +(b)	Loading & unloading at name place/port of entry in India + local transportation and storage to the consignee site + Extended Insurance for a period including 3 months beyond date of delivery**  (d)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at the Consignee's site **  (e)	

\*\* To be paid in Indian Currency (Rs.)

**# In case Full Custom duty amount not mentioned in the above format by the tenders, it will presumed that the same is inclusive in total price and nothing will be paid extra to the supplier on this account. The custom duty will reimbursed only as per SCC clause no. 1**

Total Tender price in foreign currency: \_\_\_\_\_

In words: \_\_\_\_\_

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted separately as per Section – XI – Price Schedule C
3. *The Tenderer will be fully responsible for the safe arrival of the goods at destination (consignee site) in good condition as per terms including custom clearance , payment to custom duty to custom department , insurance etc.*

Indian Agency Commission - \_\_\_% of FOB/FCA

Digital Signature of Tenderer \_\_\_\_\_

Name \_\_\_\_\_  
Business Address \_\_\_\_\_  
Digital Signature of Tenderer \_\_\_\_\_

Place: \_\_\_\_\_

Date: \_\_\_\_\_