

Amendment-I

Dated: 03.03.2017

Subject: Amendment in Offline Tender for procurement of Medical items for KCGMC, Kalpana Chawla Govt. Medical College, Karnal..

Short Tender Enquiry Document No.: HSCC/KCGMC/Medical Items/2017/ OFF/MC-03 dated 20.02.2017

1. For date/time of Submission, Closing and opening of the tender.

The closing & opening date and time now stand extended to 15.03.2017 as details below:

Sr. No	Medical Equipment for following Departments	Closing date & time for submission & receipt of Tender	Date and time of Opening of Techno – Commercial Tenders
I	Department of Physiology Equipment	15.03.2017 at 15.00	15.03.2017 at 15:30
II	Department of Pharmalogy Equipment		
III	Department of Community Medicine		
IV	Department of Biochemistry		
V	Department of Pathology Equipment		
VI	Department of Microbiology Equipment		
VII	Department of Forensic Medicine Equipment		

2. The quantity of the following item now stands amended as detailed below:

VI MICROBIOLOGY DEPARTMENT

As per Tender			Amended as	
SI No.	Name of Items	Qty (Nos)	Qty (Nos)	EMD (Rs.)
1	INCUBATORS BACTERIOLOGYCAL ELECTRICAL (LARGE) 37 degree Celsius	01	04	4000.00

2. The following item is being added to the existing list as detailed below:

VIII DEPARTMENT OF FORENSIC MEDICINE

SI No.	Name of Items	Qty (Nos)	EMD (Rs.)
07	Microtome with knife	01	7000.00

TECHNICAL SPECIFICATION OF Microtome with knife

1. High precision machine suitable for both delicate as well as hard tissue sectioning
2. Manual type with specimen clamps, user selectable coarse feed wheel turn direction, retraction on/off function and blade holder for disposable blades.
3. Section thickness setting 0.5 μ m with settings in 0.5,1,2,5 increments at different levels.
4. Section increment :
 - From 0 to 2 μ m in 0.5 μ m steps.
 - From 2 to 10 μ m in 1 μ m steps.
 - From 10 to 20 μ m in 2 μ m steps.
 - From 20 to 60 μ m in 5 μ m steps.

5. Specimen advance 28 mm or more.
6. Vertical stroke 60 mm or more.
7. Provision of step trimming.
8. Adjustable specimen clamp at least 50 x 45 mm with orientation in x y axis.
9. Single disposable blade holder for accommodating both high and low profile blades
10. Lateral coarse feed.
11. Integrated removable section waste tray.
12. Spare low and high profile blades in dispenser pack of 50 blades: 6 packets each.
13. The equipment should be USA-FDA and European–CE approved.

Accessories:-

1. Orientale specimen clamp in x y axis
2. Disposable blade holder which accommodate both high and low profile blades.
3. Spare low and high profile blades in dispenser pack of 50 blades: 6 packets each.

All other terms and condition shall remain unchanged in the aforesaid Tender.

Chief General Manager, HSCC (I) Ltd
For & on behalf of DGMER, Panchkula