HSCC (India) Ltd

<u>AMENDMENT – IV</u>

Dated: - 16.08.2016

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/ Medical Equipment / 2015-06/04 Dated 25.06.2016

Bid sale, submission and opening date for all items has been extended as per details given in Table -1:

Table -1

SI. No.	Description	Revised Schedule
i.	Sale date of the tender	23.08.2016, 2.00 P.M.
ii.	Closing date & time for receipt of tender	23.08.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	23.08.2016, 3.00 P.M.

Bidders are requested to note the following technical Amendment:

Item No. 1- MRI

Tender Specification	Amendment
Gradient System	Gradient System
b. The gradient should be actively shielded with each axis having independently	b. The gradient should be actively shielded with each axis having independently a
a slew rate of 200 T/m/s or better and a peak amplitude of 33 mT/m.	slew rate of 200 T/m/s or better and a peak amplitude of 40 mT/m.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

CGM, HSCC India Limited For and on behalf of DGMER, Panchkula