

HSCC (India) Ltd
PROCUREMENT

Amendment- IV

Dated: 05.05.2017

Subject: Invitation of Quotations for supply of Low Values Items for Physiotherapy Departments for KCGMC

Invitation for Quotation: HSCC/PUR/KCGMC/Low Value/Medical College Items/2017/ MC-Mar(4) dated 30.03.2017

a. Submission and opening date for above IFB remains as per details given below:

SI No.	Description	Schedule
1	Closing date & time for receipt of tender	16.05.2017,14.00Hrs
2	Time and date of Opening of tenders	16.05.2017,14.30Hrs
3	Venue of Opening of Tender	HSCC(I) Ltd, E-6(A), Sector-1, Noida

b. Following Items (Items under Sl. No. 29 to 32 have been inserted / included along with Technical Specifications in the above referred existing tender:

LIST OF LOW VALUE PHYSIOTHERAPY ITEMS

SL NO.	NAME OF EQUIPMENT	QTY (Nos)
1	FINGER LADDER BAR	2
2	EQUIPMENT TROLLEY	4
3	Quadriceps Bench/Table	2
4	computerized muscle stimulator	1
5	Tens Machine	2
6	Complex Exercising Therapy unit	1
7	Multi Purpose Exerciser chair	1
8	Walker with Two Wheels	1
9	Hand Exerciser Table	1
10	postural mirror	1
11	Static cycle	1
12	Almirah (full size)	4
13	WOODEN TREATMENT TABLE with mattress 4inch	10
14	Two way wooder staircase	1
15	Wrist Machine	1
16	Shoulder Wheel	1
17	Heal Exerciser	1
18	Ankle Exerciser	1
19	Grip exerciser with base	1
20	Grip exerciser without base	1
21	Stainless Steel goniometer	1
22	parallel bar	1
23	Infra Red Non Luminous Lamp	4
24	Cold Pack Unit With 12 Cold Packs	1

25	Shortwave Diathermy	2
26	Interferential Unit	1
27	Ultrasound Therapy Unit	1
28	Treatment Table With	1
	<i>Following Items have been added to the previous list:</i>	
29	Wax Bath	1
30	Hydrocollator Unit	1
31	Digital Shoulder Wheel	1
32	Infra Red Lamp	1

29. PARAFFIN WAX BATH (IMPORTED)

Approximately Specifications:

The unit should have the following features:

- The Paraffin Bath should be operatable on the “Bain Maire” principal; the paraffin is heated indirectly by the heat transferred from a liquid (water).
- Quicker heating
- More even heat distribution
- Practically no temperature fluctuations in the Paraffin.
- The wax bath should be mobile and contains a stainless steel inner tank with splash cover.
- This bath should be equipped with an electric heating element with thermostatic Temperature control
- The wax bath should be idle for dipping treatments.
- Power Consumption : 2000 Watts (Appox)
- Tank Capacity : 30 Liters.
- Heat-transfer Liquid : 10 Liters (Water).
- Temperature Range : 30 - 90 deg. C
- Safety Class : I type B IEC 601-1
- International safety standard: CE/TUV
- Internal Dimensions : 50 x 30 x 21cm (Appox)
- External Dimensions : 58 x 32 x 50cm (Appox)
- Main Voltage : 220 - 240V (50-60 Hz)

Equipment should be supplied with

- With complete accessories.
- With 50 KG paraffin wax.

30. HYDROCOLLATOR UNIT

Approximately Specifications:

- The unit should be provided with Thermostat temperature control
- The unit should have detachable insert rack to hold and suspend packs for heating.
- The unit should be supplied with twelve (12) standard Enno moist Packs of size 25 x 30 cm
- The unit should be made of stainless steel with rubber wheels for easy mobility
- The unit should be of the International Safety standard (Imported).
- The unit should be supplied with stainless steel storage rack attachment
- The unit should have energy efficient insulated casing

Equipment should be supplied with

- With complete accessories.

31. DIGITAL SHOULDER WHEEL

Approximately Specifications:

- It should have 5 functional LCD display.
- Should have magnetic resistance provides bi-directional resistive exercise.
- Should have Height adjustable wall mount allows for sitting and standing.
- Arm adjustable from 13” to 22”.
- International safety standard CE

32. INRA RED LAMP

Approximately Specifications:

- It should consist minimum 5 IR Radiators.
- It should have wavelength approx. 1000 nm.
- It should be portable & the weight should be less than 12 KG.
- It should have power of minimum 150 W for each radiators.
- It should have time adjustment upto 30 minutes.
- It should work on 230 V / 50 Hz.
- The unit should be provided with height adjustable mounting stand for various applications on the patient in sitting, standing, on supine / lying positions.
- It should have IEC / International safety standard / CE valid certificate.

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/CPPP website for corrigendum/amendments etc., if any, as these will notified on these portals only. No separate advertisement will published in the news paper in this regard.

General Manager,
Procurement
HSCC (India) Ltd,