

HSCC (India) Ltd

AMENDMENT – VII

Dated: - 29.04.2016

Subject: Amendment to the tender Enquiry Document

Ref: Tender Enquiry No.: HSCC/KCGMC/ Medical Equipment / 2015-16/01 Dated 22.02.2016

Bid sale, submission and opening date for item no. 1,3,4 & 9 has been extended as per details given in Table -1:

Table -1

Sl. No.	Description	Revised Schedule
i.	Sale date of the tender	09.05.2016, 2.00 P.M
ii.	Closing date & time for receipt of tender	09.05.2016, 2.30 P.M.
iii.	Time and date of Opening of Techno – Commercial Tenders	09.05.2016, 3.00 P.M.

Bidders are requested to note the following technical Amendment

Item No. 1. Flexible Endoscope, Upper GI & Colonoscope

S.No.	Description	Amendment by HSCC/ KCGMC
A	Video Gastro-scope – Should have the following:	
	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc.	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc. / Different modes for enhancement of Blood vessels &/ mucosa.
B	Video Colonoscope – Should have the following:	
	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc	Capable of producing Endoscopy images like NBI, FICE, I-Scan etc. Different modes for enhancement of Blood vessels &/ mucosa.

HSCC (India) Ltd

Item No. 3 Diagnostic Laparoscope (2 Nos.) & Operative Laparoscope (2 Nos)

S.No.	Description	Amendment by HSCC/ KCGMC
	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better.	The system should have Optical zoom to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used having focal length anything between 15 - 31 mm (2X) or better. Camera & Coupler should be one piece for better functioning to avoid seepage of solution & damage to camera while sterilization.

Item No. 4. Bronchoscope Video Flexible

S.No.	Description	Amendment by HSCC/ KCGMC
5	Outer diameter should be 6.0 mm or less	Outer diameter should be 5.8 to 6.2 mm.

Item No. 09. Ultrasound (Coloured) (Radiology - 1 No, Surgery - 2 Nos, Medicine - 1 No & Obs/Gyn - 5 Nos)

	Description	Amendment by HSCC/ KCGMC
Reference	Tender Specs	
Point 5	A display of 19" or more high resolution LED/LCD Flat Panel/High Resolution Screen with swivel and tilt facility. It should also have a touch panel on the console for ease of operation of minimum size 7"	A display of 19" or more high resolution LED/LCD Flat Panel/High Resolution Screen with swivel and tilt facility.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

s/d
CGM, HSCC India Limited
For and on behalf of DGMR, Panchkula