

AMENDMENT-XIII

Subject: Tender for “Execution including Design, Supply, Installation, Testing & Commissioning of Special services (SES) Works at 100 bedded Hospital at ESIC Siliguri-West Bengal”.

Tender No: HSCC/SES/ESIC/SILIGURI/WEST BENGAL/2024/65 Date:21.08.2024

With reference to the above cited subject work, the following Amendment may be noted, which shall be treated as a part of the contract to be uploaded along with tender/ bid:

1. The Last date of submission & opening of bids has been extended as follows:

Last date to fill/upload the tender: **upto 15:00 hrs. on 03.12.2024**
through e-Tendering

Date of Opening of bids: on 03.12.2024 at 16:00 hrs.

2. The validity of Bid Security/ Earnest Money Deposit (EMD) to be submitted by the bidders with their bid in the form of Bank Guarantee (BG) shall be considered from the original due date of bid submission i.e., from 05.09.2024.

3. Reply to Technical Queries raised by bidder.

Sr. No.	Vol./ Cl.No. / Ref.	Bidder Queries / request	HSCC Reply / to be amended
1.	Tech. Specs Vol 4 and BOQ Vol 5	As per technical specification Drying Tumbler – 30 kg- Electrically Heated. As per BOQ, Drying Tumbler with steam heated is mentioned. Please clarify, if electrically heated is required or steam heated.	As per technical specification.

2.	Tech. Specs Vol 4 and BOQ Vol 5	As per technical specification Flat Work Ironer – Chest Heated(Electrically heated) As per BOQ, flat work ironer with steam heated is mentioned. Please clarify, if electrically heated is required or steam heated.	As per technical specification.
3.	Tech. Specs Vol 4 and BOQ Vol 5	As per technical specification Washer Extractor – 30 kg- Electrically Heated. As per BOQ, washer extractor with steam heated is mentioned. Please clarify, if electrically heated is required or steam heated.	As per technical specification.
4.	Tech. Specs Vol 4 and BOQ Vol 5	As per technical specification Steam Generator -8KG / hr. As per amendment no. xi, technical specification has to be followed. Please make amendment in the BOQ to enable us to quote accordingly.	Technical specification will prevail rate in BOQ to be quoted as per technical specification.
5.	Tech. Specs Vol 4 and BOQ Vol 5	As per technical specification Vacuum Finishing Table with Iron Adjustable Height: Table Top Size: 1300 mm x 800 mm As per BOQ Vacuum Finishing Table with Iron Adjustable Height: Table Top Size : 1200 mm x 750 mm.	Vacuum Finishing Table with Iron Adjustable Height: Table Top Size : 1200 mm x 750 mm.
6.	Tech. Specs Vol 4 and BOQ Vol 5	As per Technical Specification Shelf Trolley, Capacity -50 Kg As per BOQ Shelf Trolley Capacity – 100 Kg	As per BOQ.
7.	Tech. Specs Vol 4 and BOQ Vol 5	As per Technical Specification Mobile Table – Table Top size : 1200mm x 750mm As per BOQ Mobile Table Top size – 1150mm x 750mm.	As per BOQ.
CSSD			
8.	Tech. Specs Vol 4 and BOQ Vol 5	Horizontal Rectangular Double Door Autoclave As per Technical Specification The Chamber and jacket should be insulated with Chloride	Horizontal Rectangular Double Door Autoclave The Chamber and jacket is insulated with 50 mm thick Chloride free mineral wool / rock wool which is

		<p>free mineral wool which should be covered under rigid, removable steel housings.</p> <p>It is requested to amend as The Chamber and jacket is insulated with 50 mm thick Chloride free mineral wool / rock wool which is covered under rigid, removable stainless steel outer cover.</p>	covered under rigid, removable stainless steel outer cover.
9.	Tech. Specs Vol 4 and BOQ Vol 5	<p>Horizontal Rectangular Double Door Autoclave As per Technical Specification All Steam Piping should be made of stainless steel with TIG welding joints. It is requested to amend as All Steam Piping made of stainless steel with Argon / TIG welding joints.</p>	<p>Horizontal Rectangular Double Door Autoclave</p> <p>All Steam Piping made of stainless steel with Argon / TIG welding joints.</p>
10.	Tech. Specs Vol 4 and BOQ Vol 5	<p>Horizontal Rectangular Double Door Autoclave As per Technical Specification</p> <ul style="list-style-type: none"> The Steam Generator pressure vessel should be made of AISI-316L Stainless steel and should be insulated in 50mm Chlorine free mineral wool enclosed in a rigid removable steel sheet housing. <p>It is requested to amend as</p> <ul style="list-style-type: none"> The Steam Generator pressure vessel should be made of AISI-316L Stainless steel and should be insulated in 50mm Chlorine free mineral/rock wool enclosed in a rigid removable steel sheet housing. 	<p>Horizontal Rectangular Double Door Autoclave</p> <p>The Steam Generator pressure vessel should be made of AISI-316L Stainless steel and should be insulated in 50mm Chlorine free mineral/rock wool enclosed in a rigid removable steel sheet housing.</p>
11.	Tech. Specs Vol 4 and BOQ Vol 5	<p>Double Door Washer Disinfector with Accessories As per technical specification Chamber Capacity Operational Volume should be 300 L.</p> <p>It is requested to amend</p> <p>Chamber Capacity: Operational Volume 285- 300 Litre.</p>	<p>Double Door Washer Disinfector with Accessories</p> <p>Chamber Capacity: Operational Volume 285- 300 Litre.</p>

12.	Tech. Specs Vol 4 and BOQ Vol 5	Heat Sealing m/c, Multi roll dispenser, Documentation Labeller, Wire Storage Shelf Module for dirty/clean/sterile area, pass box, closed transport trolley from sterile store to OT, Modular Sterilizing basket big, Modular Sterilizing basket medium, basket rack mentioned in technical specification not mentioned in BOQ.	Deleted
13.	Tech. Specs Vol 4 and BOQ Vol 5	Staff chair, lab stool without backrest, storage cupboard, waste bin pedal operated, change locker – 4 compartments, visitor chair, open storage rack, office table, shoe rack, paper dispensing trolley mentioned in specification, not mentioned in BOQ.	Deleted.
	MGPS		
14.	Tech. Specs Vol 4 and BOQ Vol 5	Supply of O2 Cylinder Class D type mentioned in technical specification not in BOQ.	Deleted.

All other terms & Conditions of the Tender shall remain unchanged.

The above amendment shall be treated as integral part of the tender document and to be submitted duly signed & stamp along with tender/bid.

Prospective bidders are advised to regularly visit through HSCC e-tender portal <https://hsc.enivida.com> & HSCC website <http://www.hsccltd.co.in> as corrigendum/amendments etc., if any, will be notified on this portal only and separate advertisement will not be made for this.

(- Sd -)
General Manager (Procurement)
HSCC (India) Ltd.