HSCC (India) Ltd

<u>AMENDMENT – XIX</u>

Subject: Amendment to the tender Enquiry Document.

Ref: Tender Enquiry No.: HSCC/KCGMC/Medical Equipment/2015-06/05 Dated 09/07/2016

Bid sale, submission and opening date for item no.8 has been extended as per details given in Table -1:

Table -1

SI. No.	Description	Revised Schedule
i.	Sale date of the tender	9.12.2016, 3.00 P.M.
ii.	Closing date & time for receipt of tender	9.12.2016, 3.30 P.M.
iii.	Time and date of Opening of Tenders	9.12.2016, 4.00 P.M.

Bidder are also requested to note the following technical amendment in the Item no.8 Sl. No. - 8 FESS Instrument

Tender Specification	Amended as
Full HD Digital Camera	
The system should have following Features:	
CCD sensing chip should optimizes image quality & Digital Source Sampling for maximizing hifidelity image transmission.	3 CCD Sensing chip should optimizes image quality & digital source sampling for maximizing hi - fidelity image transmission
Optimizes to Any Size: The system should have integrated Optical Zoom (f= 14- 30 mm, 2X) to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used. HDTV widescreen Monitor The monitor should	Optimizes to Any Size: The system should have integrated Optical Zoom (f= 14- 32 mm or more 2X) to enhance the quality of Image size & cross specialty standardization of the camera system, regardless of the telescope used.
have: Max. Resolution of 1920X1200 pixels.	Max. resolution of 1920 x 1080 pixels or better
Fiber Optic Light Cable	
Endoscopes	Endoscopes - High definition

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPPP website for corrigendum /amendments etc. if any, as these will be notified on these portals only. No separate advertisement will published in the news papers in this regard.

CGM, HSCC India Limited For and on behalf of DGMER, Panchkula

Dated: - 2.12.2016