

26	Vacuum System	
	Provide vibration mounting as per NFPA 99/HTM2022/EN737	As per Sl. No.-11
	The system shall have UL listed / CE marked control panel	As per Sl. No.-11
	Page 40	
27	As per ISO	As per Sl. No.-11
28		Tender terms prevail/ As per Sl. No.-11
29	Air compressor	
	The system shall have UL listed / CE marked control panel	As per Sl. No.-11
	Digital Dew point and CO monitor with alarm	As per Sl. No.-11
30		Tender terms prevail/ As per Sl. No.-11
	Page 42	
31	Distribution Piping	
	Painting	
	All exposed pipes should be	The painting colour of the Copper pipe shall be kept as per ISO/ IS:2379 of 1963/As per Sl. No.11
32	Alarm System 0-250 psi	Suitable psi for smooth function of Alarm system and also as per Sl. No.11
	Page 44	
33	Horizontal BedHeadPanels	

M/S
 12/9/14
 12/9/2014
 6

	Nurse Call System Module	Provision for Nurse Call system should be made in the BHP.
	Infusion Pump Mounts	Standard size
	Monitor tray with slider	Tray should be suitable for fitting/keeping of ICU Monitor
	Utility Basket	Standard size
		Tender terms prevail
34	Gas Outlet	
a	165mm long	Gas outlet must have matching adaptor and both should be imported and as per Sl. No.-11
c	25mm variation	As per Sl. No.-11
35	Valve Boxes	
b	18 gauge steel	As per Sl. No.-11
	Page 45	
c	Door frame shall	As per Sl. No.11
f	Valves shall be	As per Sl. No.-11
g	25 mm.	As per Sl. No.-11
i	51 mm	As per Sl. No.-11

M. J. S.
 12/9/14
 7 12/9/14

36	Anesthesia Gas Scavenging System	
	The duplex medical vacuum system	As per Sl. No.-11
	comply with latest edition of	As per Sl. No.-11
	One pump working and one pump will be standby.	As per Sl. No.-11
	The package should	As per Sl. No.-11
	Vacuum Pump	As per Sl. No.-11
	Page 46	
37	Control System	As per Sl. No.-11
	Electrical Distribution Panel	Tender terms prevail/As per sl. No.11
38	Accessories	
	Flowmeter with Humidifier	
	Page 47	
	Stainless steel wire mesh	As per Sl. No.-11/ As per Sl. No.-34/As per Sl No-4,5 &6

Maja
 12/9/14
 12/9/2014

	Ward Vacuum Unit	As per Sl. No.-11/ As per Sl. No.-34/As per Sl No-4,5 &6
	Theatre Vacuum Unit	As per Sl. No.-11/ As per Sl No-4,5 &6
	A3-way valve	
	Allthe above..... aluminum	As per Sl. No.-11/ As per Sl. No.-34/As per Sl No-4,5 &6
39	Page 48 onwards	As per Sl. No.-11
40		Deleted.
41		As per Sl. No.-11
42		As per Sl. No.-11
43		As per Sl. No.-11
44		As per Sl. No.-11
45		As per Sl. No.-11
46		As per Sl. No.-11
47		As per Sl. No.-11
48		As per Sl. No.-11
49		As per Sl. No.-4
50		As per Sl. No.-6
51	Standards All imported items should be of international standard like NFPA 99(latest version) standard and UL listed/CE marked or HTM 02-01	As per Sl. No.-11

MJS
12/9/14
12/9/2014

	(latest version) standard and CE marked. The entire system should be of any one standard only. All indigenous items should be of highest quality to meet the international standard and compatible to the main system.	
52	No indication relating to the imported documents in the tender document	For supply of imported items, the contractor shall submit Bill of Entry/Certificate from the manufacturer that the supplied product is brand new with Sl. No, Country of Origin and date of manufacturing along with the bill.

10/19
 12/9/14. *And*
 12/9/2014