

IFB: HSCC/PUR/PALI/01 dated. 01.09.2017
Amendment -II dated 16.10.2017

All bidders are requested to note the following:

A. Reply on Pre bid queries raised by bidder during pre bid meeting on 20.09.2017

Item No	Tender Specification	Bidder Suggestion / Request for Amendment	Amendment
Item No 5 Department Physiographs & Polygraph	Channels Student Digital Physiograph with Accessories Bio- Amplifier:- 3 Channels, Range:-+ 20mV to 100uV3	Bidder has requested to amend this parameter as "Bio-Amplifier:- 2 Channels, Range:-+ 20mV to 100uV"	As per tender Specification
	Physiograph, Single Channel with Complete Accessories 3- Channels student digital physiograph with accessories and Multi Channel Digital Polygraph System mentioned as separate items before and after Item No. 6	Bidder has requested to amend this parameter as "5a:- 3- Channels student digital physiograph with accessories and 5b:- Multi Channel Digital Polygraph System	As per tender Specification

All other terms and conditions of the tender enquiry document shall remain unchanged.

Prospective bidders are advised to regularly visit HSCC website/ CPP as corrigendum /amendments etc. if any, will be notified on this portal only, no separate advertisement will be published in the news papers.

s/d

Additional, Director

Medical Education Department, Rajasthan