

**AMENDMENT – IV****Dated: 15.01.2016**

**INVITATION FOR BIDS FOR SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF  
MEDICAL EQUIPMENTS FOR 200 BEDDED  
NEPAL- BHARAT MAITRI EMERGENCY AND TRAUMA CENTRE (NBMETC),  
KATHMANDU, NEPAL**

**BID REFERENCE: IFB No. HSCC/PUR/NEPAL/2015-16/1 Dated 10.11.15**

Bidder is requested to note the following amendment technical specification of item no.06:

**Section – IV Technical Specifications**

**Item No. 6 C-ARM**

Description	Amended as
<p><b>1. Generator &amp; X-ray Tube</b> Generator and X-ray Tube Generator should be high frequency type, with power output of 2.5KW or more....</p> <p>The Generator should be capable of providing Pulse Fluoroscopy with Pulse rates upto 20 frames/sec or more.</p> <p>6. Heat storage capacity should be 35KJ or more. Minimum of 20 minutes of continuous fluoroscope should be possible</p>	<p><b>1. Generator &amp; X-ray Tube</b> Generator and X-ray Tube Generator should be high frequency type, with power output of <b>2.0 KW</b> or more.....</p> <p>The Generator should be capable of providing Pulse Fluoroscopy with upto 20 frames/sec or more.</p> <p>6. Heat storage capacity should be 35KJ or more/ apporx 46 KHU. Minimum of 20 minutes of continuous fluoroscope should be possible</p>
<p><b>2. C-Arm:</b> Fully counterbalanced C-Arm movement with integrated cables and electromagnetic brakes. Give details of:</p> <ul style="list-style-type: none"> <li>• Source to detector distance</li> </ul>	<p><b>2. C-Arm:</b> Fully counterbalanced C-Arm movement with integrated cables and brakes. Give details of:</p> <ul style="list-style-type: none"> <li>• Source <b>to image</b> distance</li> </ul>
<p><b>4. Image viewing:</b> System should have a hard disc drive storage capacity of 200GB or more with post-processing facility</p>	<p><b>4. Image viewing:</b> System should have a hard disc drive storage capacity of 200GB/20000 images or more with post-processing facility</p>
<p>5. Image Documentation: Essential Accessories Sterile covers (100 disposables) for C-Arm, X-ray tube and Flat Panel Detector.</p> <p><b>IMPORTANT INSTRUCTIONS:</b> e. DICOM &amp; PACS feature</p>	<p>5. Image Documentation: Essential Accessories Sterile covers (100 disposables) for C-Arm, X-ray tube and <b>Image intensifier</b></p> <p><b>IMPORTANT INSTRUCTIONS:</b> e. DICOM &amp; PACS option should be available</p>

All other terms and conditions of the bid document shall remain unchanged.

S/d  
Joint Secretary –DPA (III)  
Ministry of External Affairs, New Delhi