

Dated: 19.01.16

AMENDMENT No – VII

Subject: Amendment to the tender Enquiry Document

Ref: Tender Enquiry No.: HSCC/PUR/MEA - Myanmar/ Equipment/2015-16 dated 09.10.2015

Bidders are requested to note the following amendment:

Section – VIII Technical Specifications

Package 9 Auto Analyzer

Tender Point No.	Tender Specification	Amended/ Read As
Point 3.1	<ul style="list-style-type: none">• Optical Requirement Wavelength Range: 340 to 700nm• Flow cell volume: approx. 50µl	<ul style="list-style-type: none">• Optical Requirement Wavelength Range: 340 to 650nm• Flow cell volume: approx. 50µl/ Cuvette based/latest technology
Point 3.5	Inbuilt printer thermal type with 40 characters/line or better	Inbuilt printer thermal type with 40 characters/line or better alternatively external printer

All other terms and conditions of the bid document shall remain unchanged.

S/d
Joint Secretary –DPA (III)
Ministry of External Affairs,
New Delhi