

Amendment – III

Project: Procurement of MRI Machine for Yangon Children Hospital, Yangon, Myanmar

Bid No.: IFB Reference: HSCC/PUR/MEA - Myanmar/Equipment - MRI/2014 dated 20th May 2014

Bidder are requested to note the following clarification/ amendment after the pre bid meeting held on 03.06.2014

Commercial conditions:

S. No.	Existing Bid Clause	Clarification / Amended as
1.	GCC Clause 7.6, 12.3 (9), 12.4(7) : Inspection certificate for the dispatched equipments issued by recognized/ reputed agency like SGS, Lloyd or equivalent (acceptable to the Purchaser) prior to dispatch	Amended as GCC Clause 7.6, 12.3 (9), 12.4(7): "Inspection certificate issued by recognized/ reputed agency (ISO certified Inspection agencies) like SGS, Lloyd or equivalent prior to dispatch certifying quantity, quality, packing and marking "
2.	GCC Clause 24 Taxes and Duties Supplier shall be entirely responsible for all taxes , duties, license fees etc incurred until delivery of the contract goods to the purchase	Clarification Clause 6 (a) & (b) of ITB indicates that custom clearance will be arranged by Government of Myanmar, hence no custom duty is to be paid as the clearance is to be arranged by Government of Myanmar.
3.	Contact person at Children Hospital Yangon	Clarification The details of the contact person is : Prof Htoo Han , Medical Superintendent (Acting) Yangon Children Hospital, Yangon Ph. No. +9595124112 drhatoohan@gmail.com

Technical Specification:

S. No.	Existing Tender Specification	Amended as
1.	3 Patient comfort features i. Patient monitoring devices for ECG, respiratory, pulse rate, oxygen saturation, at the console etc. A comprehensive solution at patient side and at main console capable of gating the sequence protocols with respect to patient's heart and respiratory rates. Remote display of gating signals on magnet and at console.	3 Patient comfort features i. Patient monitoring devices for ECG, respiratory, pulse rate, oxygen saturation, at the console etc. A comprehensive solution at patient side and at main console capable of gating the sequence protocols with respect to patient's heart and respiratory rates. Remote display of gating signals at console.

S. No.	Existing Tender Specification	Amended as
2.	xxii. Sequence for in-line motion correction for uncooperative patients/ children (with software and acquisition – sequences like BLADE, PROPELLAR, k-t BLAST or equivalent)	xxii. Sequence for in-line motion correction for uncooperative patients/ children (with software and acquisition – sequences like BLADE, PROPELLAR, k-t BLAST, Multivane XD or equivalent)(latest release).
3.	<p>Accessories:</p> <p>i. Hand held metal detectors (2 Nos.)</p> <p>ii. MR compatible (min. 2000 Gauss line limit) cardiac and physiological monitor (ECG, NIBP, SP02) with both pediatric & adult probes with MR compatible ECG electrodes (2000 in number).</p> <p>iii. MR compatible (min. 1000 Gauss line limit) syringe/infusion pump.</p> <p>iv. MR Compatible Dual Pressure injector (min. 2000 Gauss line)with 300 spare syringes & tubings.</p> <p>v. MR compatible anaesthetic system (for Paediatric and adults use) with dual vaporisers (for isoflurane, halothane) CE (Europe)/ FDA approved, and other accessories including MR compatible patient trolley, gas cylinder (2 each for N2O and O2) , I/V stand, pediatric incubator and laryngoscope.</p>	<p>Accessories:</p> <p>i. Hand held metal detectors (2 Nos.)</p> <p>ii. MR compatible (min. 2000 Gauss line limit) cardiac and physiological monitor (ECG, NIBP, SP02) with both pediatric & adult probes with MR compatible ECG electrodes (2000 in number).</p> <p>iii. MR compatible (min. 1000 Gauss line limit) syringe/infusion pump.</p> <p>iv. MR Compatible Dual Pressure injector (min. 2000 Gauss line) with 300 spare syringes & tubings.</p> <p>v. MR compatible anaesthetic system (for Paediatric and adults use) with dual vaporisers (for isoflurane, halothane) CE (Europe)/ FDA approved, and other accessories including MR compatible patient trolley, gas cylinder (2 each for N2O and O2) , I/V stand, pediatric incubator and laryngoscope.</p> <p>vi Suitable UPS for installation and day to day running of the required MRI system with 30 min back up.</p> <p>Bidder to quote generator as a part of optional item.</p>

All other terms and conditions shall remain unchanged.

S/d
Joint Secretary –DPA (III)
Ministry of External Affairs,
JNB, Janpath, New Delhi