

To,
All Bidders

Amendment-I

Subject: Tender (Third Call) for Supply, installation, testing & commissioning of Hospital Furniture & other related miscellaneous works like minor civil, electrical, PHE works etc. for CCU at AIIMS, Ansari Nagar, New Delhi

Dear Sir,

This has reference to the pre bid meeting held on 27.05.2014 at 12:00 hrs at HSCC corporate office Noida for the subject works,

The following AMMENDMENT-I/CLARIFICATION-I may be noted, which shall be treated as a part of the contract to be submitted along with tender.

Sl no	Queries raised by bidders during pre-bid meeting	Clarifications
1.	Item Name: <u>Bed Side Locker</u> : You have asked for two lockable castors provided for easy handling. We request for all 4 castors to be without lockable because usually locker does not required lockable castors, because it is not easy to lock & unlock, due to castor size being very small (i.e. 50 mm)	No change, as per tender conditions only
2	Item Name: <u>Over bed Adjustable Table</u> : In first column you have asked for food table is an over bed table for patient height adjustable by Gas Lift. In 5 th column you are asking for Plastic handle with SS rod insert for a strong and provide the wide grip. Smooth functioning Crank handle with adjustable height and consistent motion during operation. Please clarify: Whether you want Gas operating beside table or Crank handle for adjustable of height beside table shall be provided with gas lift or with Crank handle. Because both the system cannot be incorporate in single product of bed side table.	Gas lift over bed Adjustable Table may be provided in place of Crank Mechanism
3.	<u>Item Name: Foam Mattress</u> in 3 rd column you are asking for SSPT rail system provided solid edge support allowing easy patient transfer. We request you allow to offer our standard mattress. Mattress does not require any railing system.	Mattress without rail system may be provided

Other all terms & conditions of tender shall remain unchanged.

Prospective bidders are advised to regularly scan through HSCC web site as corrigendum/amendments etc., if any, will be notified on HSCC web site and separate advertisement will not be made for this.