

INVITATION FOR QUOTATION ONLINE

HSCC/PUR/PALI/ Community Medicine/ Quot/02

Dated: 6th Nov. 2017

To

All Bidders

Subject: Invitation **on line** for Quotations for supply of Low Equipment for **Community Medicine** for Medical College, at Pali, Rajasthan.

Dear Sirs,

1. HSCC (India) Ltd. for and on behalf of Medical College, at Pali, Rajasthan. invites your **on line** most competitive quotation for the following goods of the respective departments quantity mentioned the **section I** and technical specification mentioned in the **section II**.
2. Quotation:
 - 2.1 The contract shall be for the full quantity as described above.
 - 2.2 Corrections, if any, shall be made by crossing out, initialling, dating and rewriting.
 - 2.3 The **on line** prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
 - 2.4 The unit price/ rate of the item should be clearly indicated in the quotation. Rates/Prices quoted shall be including of forwarding, and insurance & transportation, warranty, but **excluding GST** up to the consignee **Medical College, at Pali, Rajasthan**.
3. Each bidder may quote one item or more than one item as per **section III (Price format) with equipment technical literature**.
4. Evaluation of Quotations:

The Purchaser shall evaluate and compare the quotations determined to be substantially responsive i.e. which;

 - 4.1 are properly signed; and
 - 4.2 confirm to the terms and conditions, and specifications.
 - 4.3 final considerations of equipments shall be based on the quality of equipments during demonstration / inspection.
 - 4.4 The quotation will be evaluated and compared separately for each item.
 - 4.5 The Purchaser reserves the right to ask for a free demonstration/ sample approval of the quoted equipment at a pre determined place acceptable to the purchaser for technical acceptability as per the tender specifications, failing which bid may not be consider.
5. The Quotations would be evaluated item wise.
6. Award of contract:

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest rate for the item subject to quality of the items during demonstration / inspection.

 - 6.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and cancel the bidding process and reject all quotations at any time prior to the issue of Purchase Order/contract, without assigning any reason.
 - 6.2 At the time of awarding the contract, the purchaser reserves the right to increase or decrease by up to twenty five (25) per cent, the quantity of goods and services mentioned in the schedule (s) in the "List of Requirements" (rounded of to next whole number) without any change in the unit price and other terms & conditions quoted by the tenderer.
7. Delivery period shall be within **10 days** from date of placement of order at Consignee i.e. **Medical College, at Pali, Rajasthan**.

8. Payment shall be made only in Indian Rupees as follows:

Satisfactory Acceptance and delivery - 100% of total cost

100% Payment on submission of following documents (Duly signed & stamped at your end):-

- Copy of Purchase order.
- Final acceptance certificate issued by Client/HSCC
- Invoice in favour of consignee/MEDICAL COLLEGE, AT PALI, RAJASTHAN through HSCC (I) Ltd
- Warranty Certificate in original.

9. All supplied items shall be under **one year warranty** from the date of successful acceptance by **MEDICAL COLLEGE, AT PALI, RAJASTHAN**.

10. You are requested to provide your lowest offer as follows:

Sr. No.	Item no	Name of items	Closing date & time for submission	Date and time of Opening of Techno
1	Department of Community Medicine	List attached	13 th Nov. 2017 at 14:00	13 th Nov. 2017 at 14:30

NOTE: (Under any unforeseen circumstances if the due date for submission of Tender is declared as holiday then the tender shall be submitted & opened on the next working day at the scheduled time).

The quotations will be opened in HSCC office, NOIDA as mentioned above in the presence of bidders or their authorized agents as they may choose to attend

11. Information brochures/ Product catalogue, if any, must be accompanied with the quotation clearly indicating the model quoted for.
12. Sealed quotation to be submitted/ delivered at the address mentioned below:

On line submission: Quotation Form duly signed & filled and Technical compliance sheet with technical leaflets and Price Bid /Financial Bid

**General Manager (Projects) HSCC
(India) Ltd.
E- 6 (A), Sector -1.NOIDA -201 301.**

13. Quoted amount should be in Indian Rupees only and as per **section -III** price schedule.
14. For all items, the Technical Evaluation Committee may opt for Demonstration of the items. The Committee may also ask for Demonstration / Inspection before supply / delivery of the items for quality assurance.

Note: Please indicate the quotation reference no. (given at the top of page 1 of this letter) and Serial No of the Items on your offer.

The details of various medical equipments shall be also made available at www.hsccltd.co.in, and modification/amendments etc, if any, shall only be notified on website only.

We look forward to receiving your quotation and thank you for your interest in this project.

General Manager (Projects) on
behalf of Additional Director
Medical Education, Jaipur, Rajasthan

Section -I

LIST OF ITEMS & QUANTITIES

Sr. No.	Equipment	Quantity	Warranty from date of Acceptance
1	Museum jars	25	1 Year
2	Baby weighing machine	2	
3	Refrigerator 9 Cu. Ft.	3+ Additional one each at RHTC and UHTC	
4	Ice Lined Refrigerator (I.L.R.) (at Health Centre)	1	
5	Public Address system (2 portable for field based activities and one each for RHTC & UHTC)	4Sets	
6	MUAC tapes	10	
7	GlobeThermometer	3	
8	Anemometer	3	
9	Sound level meter	3	
10	Soil testing kit	1	
11	Water sampling bottle from any depth	1	
12	Needle Shredder	3	
13	Vaccine carrier	5	
14	Craft water testing kit	1	
15	Protective devices for occupational safety	3sets	
16	Ear muffs	3sets	
17	Ear plugs	3sets	
18	Safety helmet	3sets	
19	Goggles	3sets	
20	Safety boots	3sets	
21	Swine flu kit	3sets	
22	Gloves	3sets	
23	Triple layer surgical mask	3sets	
24	High efficiency mask	3sets	
25	Long sleeved cuffed gown	3sets	
26	Protective eye wear	3sets	
27	Cap	3sets	
28	Disposable delivery kit	3sets	
29	Treatment kits as per national health programs	3each	
30	Iodine testing kit	10	
31	Glucometer	10	
32	Slide set for entomology	5	

33	Mosquito catching kit	3
34	Clinical Thermometer	10
35	Solar radiation thermometer	3
36	Treatment kits as per National Health Programs	3each
37	Iodine testing kit	10
38	Glucometer	10
39	Slide set for entomology	5
40	First Aid Kit	1
41	Spirometer	3
42	Audiometry	1
43	Otoscope	1
44	Ophthalmoscope	1

SectionII

Technical Specification of Equipments

1	Museum jars	25
2	Baby weighing machine	2
3	Refrigerator 9 Cu. Ft.	3+ Additional one each at RHTC and UHTC
4	Ice Lined Refrigerator (I.L.R.) (at Health Centre)	1

1. Description of Function:

- 1.1 Ice lined refrigerators maintain temperature of +2°C to + 8°C. Not more than 8 h'l'S continues or intermittent power should be sufficient per 24 hrs to maintain vaccine temperature below 8°C.
- 1.2 Ice-lined refrigerators are required at district and regional levels, since electricity suppliers are rarely perfect and standby electricity suppliers may not be available.

2. Operational Requirements:

- 2.1 Designed for tropical climates.
- 2.2 Target holdover time should be 20 h'l'Sor more in a continues external temperature of 43°C.
- 2.3 Hot and Cold compressor starting at 172 Volts (22% below rated voltage).
- 2.4 Manufacturing process of the product should not use or produce hazardous chemical gases.
- 2.5 Provision for drainage for the waste water.
- 2.6 Should have legs in the base with rotating screw type height adjustments to balance the weight all uneven floor.
- 2.7 The unit should have ground clearance of minimum 100 mm.

3. Technical Specifications:

- 3.1 Net Storage Capacity: 250-300 liters within basket in place.
- 3.2 Construction:
 - (a) Internal.-Stainlesa 304 grade steel/Pre Painted Galvanized.
 - (b) An additional special ice lining consisting of ice packs covered by strong plastic shell.
- 3.3 External: Corrosion Resistance.

- 3.4 Chest type with CFC-free insulation.
- 3.5 Should have horizontal water cool pack covering the top of the basket.
- 3.6 Solid door with lock and handle
- 3.7 Type: Compression Cycled, CFC free (both for refrigeration and insulation). All system tubing (suction tube, freezer tube and condensing tube) should be of minimum 99.97% of copper coil.
- 3.8 Temperature of a full vaccine and medicine to remain +2 deg C to +8deg C during continuous availability. of energy at ambient temperature +5deg C to +45 deg C with intermittent! continues electricity supply 8 hrs in a 24 hrs cycle. The temperature difference between any two points in the cabinet should not be more than +2 deg C once stabilized.
- 3.9 _Inlet of Capillary should he outside the PUF body.
- 3.10 ON/OFF Switch and power indicator should be available.
- 3.11 A Micro Processor based control unit should be provided for setting of temperature and display following features:
 - 3.11.1 3 digit digital display (to one decimal point) of cabinet temperature. The sensor should be placed 25 to 50 mm above base of storage chamber.
 - 3.11.2 Power on LED/LCD indicator
 - 3.11.3 Audio (minimum 65 dBA) and visual alarm against the violation of temperature ranger (less than +2 and more than +8 degree C)
 - 3.11.4 Min. & max. cabinet temperature digital display of 24 hrs. and breaches during last 24 hrs.
 - 3.11.5 The unit should be sealed/protected from dust, moisture or condensed water falling over it.
- 3.12 Accuracy for digital controller +0.5 degree centigrade.
- 4 System Configuration
 - 4.2 Programmable Micro-processor control unit child lock facility.
 - 4.2 .Shouldhave provision to set minimum and maximum temperature at 0.1 degree Centigrade to programme the unit for continuous operation.
 - 4.3 Should have provision for defrosting program.
5. Accessories and spares
 - 5.1 Food Storage Basket allowing free circulation of ail', having the size to be able to accommodate 4 to 6 of them in them in the unit and suitable to match the net volume requirement. It should be minimum 5 wire basket.
 - 5.2 Stem.Alcohol.thermometer (specification and standard as per MOHFW approved Annexure-I) - one piece per unit range of -30 to+ 50 degree.centigrade.
 - 5.3 The supplier is required to maintain the entire spare throughout the guarantee period.
6. Environmental factors:
 - 6.1 The unitshall be capable ofbeing stored continuously in ambient temperature of 0 to 50deg C and relative humidity of95%
 - 6.2 The.unit shall'be.capable.orQP~ra:ting.continuously in ambient temperature of 5 to 45 deg C and relative humidity of 90% ..
 - 6.3 The plug should be flexible and unbreakable sealed rubber type.
7. Power Supply
 - 7.1 Power input to be 220-240 V AC , 50 Hz as appropriate fitted with Indian Plug.
 - 7.2 Suitable capacity Voltage Stabilizer as per requirement ofLLR.
8. Standards & Safety
 - 8.1 Product Should be CE Approved
9. Documentation:
 - (i) A paper copy of use fiioperator manuals tobe supplied in English.
 - (ii) .A paper. copy of technical/wiring 'diagram/maintenance manuals to be supplied in English.
 - (iii) Certificate of Testing should be submitted at the time of supply from NABL certified lab / ILAC/ STQC Labs.
10. Packing of the equipment during shipment:
 - (i) The supplier should provide strong and sufficient packing to ensure safe arrival of goods at the destination free from loss or damage.
 - (ii) Vertical arrow should be marked at all sides of packages to ensure

Terms & Conditions

1. Guaran tee: Two years on equipment from the date of installation.
2. CMC: CMC will be given@5% (of net rate-inclusive of Excise Duty & exclusive ofVAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of guarantee period of two years.
3. Installation and Training will be done by the supplier free of cost.
4. The service engineer should be based in Rajasthan.

5. The company should mention the make and model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical compliance sheet as per technical specifications.

5	Public Address system (2 portable for field based activities and one each for RHTC & UHTC)	4Sets
6	MUAC tapes	10
7	Globe Thermometer	3
8	Anemometer	3
9	Sound level meter	3
10	Soil testing kit	1
11	Water sampling bottle from any depth	1
12	Needle Shredder	3
13	Vaccine carrier	5
14	Craft water testing kit	1
15	Protective devices for occupational safety	3sets
16	Ear muffs	3sets
17	Ear plugs	3sets
18	Safety helmet	3sets
19	Goggles	3sets
20	Safety boots	3sets
21	Swine flu kit	3sets
22	Gloves	3sets
23	Triple layer surgical mask	3sets
24	High efficiency mask	3sets
25	Long sleeved cuffed gown	3sets
26	Protective eye wear	3sets
27	Cap	3sets
28	Disposabledeliverykit	3sets
29	Treatmentkitsaspernationalhealthprograms	3each
30	Iodine testing kit	10
31	Glucometer	10
32	Slide set for entomology	5
33	Mosquito catching kit	3
34	Clinical Thermometer	10
35	Solar radiation thermometer	3
36	Treatment kits as per National Health Programs	3each
37	Iodine testing kit	10
38	Glucometer	10
39	Slide set for entomology	5
40	First Aid Kit	1
41	Spirometer	3
42	Audiometry	1

43	Otoscope	1
44	Ophthalmoscope	1
45	Band Saw	1 for dept. of Anatomy

Section-III

PriceBid

Sr no.	Name of item	Qty	Unit cost (Rs.)	GST/ Sales Tax /service tax		Unit cost included GST/ Sales Tax /service tax (Rs.)	Total cost included GST/ Sales Tax /service tax (Rs.)
				%	Amount (Rs.)		
				a	b		
1							
2							

Note:

1. Total cost of all items shall be included of all packing & forwarding, freight charges & insurance from ware house to consignee at **Medical College Pali**, local tax, entry tax, duties, **one year warranty**, VAT/ taxes /GST and other levies payable by the supplier under the contract.
2. Bidder shall fill all cost i.e. a,b,c... failing which it will presumed that the same is inclusive in the total price and nothing will be paid on this account extra.
3. Bidder shall mention present rate of GST, failing which it will be presumed that the same is inclusive in the total price and nothing will be paid on this account extra.
4. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail

**FinalAcceptance Certificate Handing
over]**
**(To be given by consignee's authorized
representative)**

The following store (s) has/have been installed & commissioned in good working satisfactory condition:

1. Contract No. & date :
2. Supplier's Name :
3. Consignee's Name & Address :
4. Name of the item supplied :
5. Date of Handed over to consignee :
6. Name consignee /HSCC Representative :
7. Signature of consignee/HSCC Representative :
8. Seal of the Consignee :